

Phone: (520) 741-1404
 Fax: (520) 741-0762
 email: info@dataforth.com

TEST DATA SHEET

ACCURACY TEST

ed) ==	Measured Vout (V) * =====	Error (%) =====	Status =====
	-9.903	+0.003	PASS
	-5.001	-0.003	PASS
	-0.000	-0.000	PASS
	+4.999	-0.003	PASS
	+9.996	-0.003	PASS

FINAL TEST RESULTS

Measured Value =====	Specification =====	Status =====
24.7 mA	< 50 mA	PASS
20 ohms	< 55 ohms	PASS
0.003 %	+/- .03 %	PASS
0.003 %	+/- .08 %	PASS
426 uV/%	+/- 1200 uV/%	PASS
50.6 dB	61+/-16 dB	PASS
100.5 %	85 to 110 %	PASS
127 uVrms	< 600 uVrms	PASS
		PASS
		PASS

Check List

Mounting Screw: 

Module Header: 

QC: 

that the above product is in conformance with
 extent specified. This product is not
 for use in life support devices and/or systems.