

## DSG-HDICE Meeting 11-April-2019

**Date: 04/11/19**

**Time: 2:00pm**

*Attendees: Mary Ann Antonioli, Peter Bonneau, Amanda Hoebel, Xiangdong Wei*

### 1. Windows 10 Upgrade:

- 1.1. Service request was submitted for Windows 10 PC not updating to update version 1908.
  - 1.1.1. Computer Center said that the 1908 update was not approved by the lab.
    - 1.1.1.1. The lab should approve the update in a few weeks.
    - 1.1.1.2. When the lab approves the update, HDICE will try to update the PC again.

### 2. IBC Magnet Repair:

- 2.1. Machine shop is working on rewinding the coil to the magnet.
  - 2.1.1. Coil rewinding will take 6 weeks to complete.

### 3. Dilution Fridge Debugging:

- 3.1. DSG will take a look at the Dilution Fridge card tomorrow, 4/12/19.
  - 3.1.1. Suggested to monitor voltages at specific check points to record the voltage when the power supply trips.

### 4. Schedule:

- 4.1. HDICE is planning for a summer beam run in UITS.
  - 4.1.1. Several reviews need to be conducted to approve run.