

Hall D Detector Hall D Time-of-Flight Checklist

Document ID: 465

Group: Hall D Detector

Subsystem: Hall D Time-of-Flight

Author: Ito, Mark (marki)

Submitted By: Ito, Mark (marki)

Revision Number: 1

Revision Date: 2014-06-17 18:14

Revision Comment: first attempt

1. check that high voltage is on
2. check that the EPICS HV alarm handler is running
3. check that all EPICS alarms are clear
4. check that all counters have cosmic hits
 1. take data with the TOF cosmic trigger
 2. look at the TOF monitoring histograms and check that they are nominal